





QUANTITY	CONTENTS	CODE	
30 mL	*Silica Reagent #1	*4571-G	
15 mL	*Silica Reagent #2	*4467-E	
15 mL	*Silica Reagent #3	*4468-E	
5 mL	Reducing Reagent	6405-C	
2	Test Tubes, 5 mL, glass, w/caps	0106	
1	Pipet, plain, plastic	0352	<div><p>*Reagent is a potential health hazard. READ SDS: lamotte.com</p><p>Emergency information: Chem-Tel USA 1-800-255-3924 Int'l, call collect, 813-248-0585</p><div></div></div>
1	Pipet, 0.5 mL, plastic	0353	
1	Silica Octa-Slide 2 Bar, 0.5-10.0 ppm	4465-01	
1	Octa-Slide 2 Viewer	1101	

To order individual reagents or test kit components, use the specified code number.

Warning! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully.
Not to be used by children except under adult supervision.

USE OF THE OCTA-SLIDE 2 VIEWER

USE OF THE
OCTA-SLIDE 2 VIEWER

PROCEDURE

1. Insert the Silica Octa-Slide 2 Bar [4465-01] into the Octa-Slide 2 Viewer [1101].

2. Fill test tube [0106] to 5 mL line with sample water.

3. Add 7 drops of *Silica Reagent #1 [4571]. Cap and mix by inverting 4 times.

4. Add 6 drops of *Silica Reagent #2 [4467]. Cap and mix. Wait 5 minutes.

5. Add 6 drops of *Silica Reagent #3 [4468]. Cap and mix. Wait 2 minutes.

6. Use pipet [0352] to add 2 drops of Reducing Reagent [6405]. Cap and mix. A BLUE color will develop in 10 seconds if silica is present.

7. Insert test tube into the Octa-Slide 2 Viewer. Hold the viewer so that non-direct light enters through the back. Match sample color to a color standard. Record as ppm Silica.
- NOTE: If test color is darker than the 10.0 ppm standard, repeat test on diluted sample. Use pipet [0353] to add 0.5 mL of water sample to test tube. Dilute to 5 mL line with deionized water. Follow Steps 2 through 6 above. Multiply result by 10. Record as ppm Silica.