

# DECLARATION OF CONFORMITY

---

**Application of Council Directives:**

EMC Directive 89/336/EEC

**Standards to which Conformity Declared:**

**EMC:**

EN61326: 1997, Class B Emissions

EN61000-3-2:2000 Harmonic Current Emissions

EN61000-3-3:1995 Voltage Flctuations

EN61326: 1997, Minimum Immunity Requirements

47 CFR:2003, 15.107 AND 15.109 Class B

**SAFETY:**

IEC 61010-1/ EN61010-1 Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1 general use.

**Brand Name:**

LaMotte Co

**Manufacturer:**

Apprise Technologies Inc.

**Type of Equipment:**

Portable Turbidity, Color and Chlorine Meters

**Model Name:**

TC3000e, TC3000i, 2020e, 2020i

**Year of Manufacture:**

2005

**Testing Performed By:**

Intertek ETL SEMKO  
7250 Hudson Blvd  
Suite 100  
Oakdale, MN 55128


*I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive and Standards.*

802 Washington Ave, Chestertown, MD 21620

**Place**

5/6/2020

**Date**

  
\_\_\_\_\_  
**Signature**

James W. Egan Jr.

**Name**

VP Research and Development

**Position**