## **DECLARATION OF CONFORMITY**

Application of Council Directives:

Radio and telecommunications terminal equipment (RTTE) Directive 1999/5/EC or Radio Equipment Directive 2014/53/EU

EUCB, IEC/IECEE Safety requirements FCC Title 47 CFR Part 15 Subpart B

ICES-003 Issue 6 January 2016 AS/NZS CISPR22 2009

**Standards to which Conformity Declared:** 

#### **Emissions:**

ETSI EN 301 489.1, Section 8.4. AC Mains Power Input/Output Ports

ETSI EN 301 489.1, Section 8.5. Harmonic Current Emissions (AC Mains Input Port)

ETSI EN 301 489.1, Section 8.6. Voltage Fluctuations and Flicker (AC Mains Input Port)

#### Immunity:

ETSI EN 301 489.1, Section 9.2. Radio Frequency Electromagnetic Field

ETSI EN 301 489.1, Section 9.3. Electrostatic Discharge

ETSI EN 301 489.1, Section 9.4. Fast Transient Common Mode

ETSI EN 301 489.1, Section 9.5. Radio Frequency Common Mode

ETSI EN 301 489.1, Section 9.7. Voltage Dips and Interruptions

ETSI EN 301 489.1, Section 9.8. Surges

#### Conducted:

47 CFR Part 15.107(a). Conducted Emissions Limits for a Class B Digital Device.

ICES-003 Issue 6 January 2016. Conducted Emissions Limits for a Class B Digital Device.

AS/NZS CISPR22 2009. Standard Conducted Emissions – AC Voltage Class B

### Radiated:

47 CFR Part 15.107(a). Radiated Emissions Limits for a Class B Digital Device.

ICES-003 Issue 6 January 2016. Radiated Emissions Limits for a Class B Digital Device.

AS/NZS CISPR22 2009. Standard Radiated Emissions - Class B

# **DECLARATION OF CONFORMITY**

	Safety:
	EN 61010-1: 2010, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use. Part 1: General Requirements
	IEC 61010-1: 3rd Edition, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use. Part 1: General Requirements. Account for AS/NZs
Manufacturer's Name:	LaMotte Company
Manufacturer's Address:	802 Washington Ave PO Box 329 Chestertown, MD 21620
Type of Equipment:	Water Test Instrument
Model Name:	Waterlink Spin Touch
Year of Manufacture:	2017
Testing Performed By:	Met Laboratories, INC 914 West Patapsco Ave Baltimore, MD 21230
I, the undersigned, hereby above conforms to the above L	declare that the equipment specified Directive and Standards.
802 Washington Ave, Chestertown, MD 21620	James W Egamp
Place	Signature
2/28/2017	James W. Egan Jr.
Date	Name
	VP Research and Development
	Docition