



Code 3072-J

InstaTest[®] ANALYTIC

QAC ppm

0	100	200	300	400	500

Immerse pad in solution for 10 seconds.
SHAKE OFF EXCESS WATER.
Compare pad to color chart.

View from above.

- Sample must be at room temperature (above 75°).
- Keep out of sunlight.
- Keep vial closed when not in use.

Code 3072-J



Lamotte Company
Chestertown, MD 21620 USA
www.lamotte.com

